



BIOMEDICAL SYSTEMS
370 West 1700 South
Logan, Utah 84321 USA
435 752 6011
800 453 2725
fax 435 752 4127
web www.elitechgroup.com

2016 April 18

Product Bulletin:

Non-Hazardous Components

Eon ISE products

ISE Module Safety Data Sheets

Attention: Safety Officer / SDS Requester

Following is a list of products required to operate and maintain the Eon-series Clinical Chemistry analyzers. This list identifies which products require and which do not require a Safety Data Sheet (SDS).

According to OSHA regulations, substances containing less than 1% of a hazardous material or less than 0.1% of an extremely hazardous material do not require an SDS.

Product Description	ITEM	SDS	Contents
ISE Reagent Pack	79483	No	Saline solution, buffer, preservative
Sodium (Na) Electrode	79478	No	Saline solution, buffer
Potassium (K) Electrode	79479	No	Saline solution, buffer
Chloride (Cl) Electrode	79480	No	Saline solution, buffer
Reference (Ref) Electrode	79473	No	Saturated KCl solution